

## 附件二【技術規格核對清單】

附註：○ 為必要的配備，△ 為理想的配備；

請於下列表右方列以"有"或"沒有"填寫所對應之項目。

### 1. 必需是一枱上高效液相色譜-高分辨率質譜儀聯用儀

1.1 其必需包括High Performance Liquid Chromatograph、High-Resolution Mass Spectrometry System、System Control and Data Processing Unit、Uninterrupted power supply、Nitrogen generators及兩台電腦。

1.2 包括所提供的工作枱在內，其總高度不可超過220厘米。

○	
○	

### 2. High Performance Liquid Chromatography

#### 2.1. Solvent delivery unit

2.1.1. A binary solvent delivery design, with an internal 4-solvent selection valve included, capable of delivering isocratic and binary gradients with smooth motion control.

2.1.2. With gradient operation, the unit shall be able to deliver at least 2 solvents and programmable from 0 to 100% in 1% increment, or finer, for each of the solvents with the total composition summing up to 100%.

2.1.3. Solvent composition precision should be < 0.3%RSD.

2.1.4. Solvent composition accuracy should be ±0.35% absolute or better.

2.1.5. The unit should have a software controllable automatic or manual priming function for module fluidic path.

2.1.6. The operation flow rate shall range from 0.001 ml/min to 5mL/min or wider in 0.001 mL/min increments or finer.

2.1.7. The flow precision shall be smaller than 0.1% RSD or <0.01 min SD or better, based on retention time.

2.1.8. Flow accuracy should be ±1% or 10µL/min or greater.

2.1.9. The dead/delay/dwell volume after the LC column should be equal or <100µL.

2.1.10. The operating pressure should be up to 1000 bar.

2.1.11. Leak detection system shall be available with leak safe handing design to assure isolation of electrical components from liquid flow path. Automatic shutdown of the pumping system should be initiated when leak is detected.

2.1.12. Information for diagnostic purpose shall be available for continuous tracking of instrument usage in terms of solvent utilization status associated with user defined

2.1.13. System software should control operation of the entire unit.

2.1.14. An integrated solvent degassing with at least 2 channels.

2.1.15. The module should be able to run with LC column with sub-2µm particle size column, at flow rates 0.5 mL/min without any hardware modifications.

○	
○	
○	
○	
○	
○	
○	
○	
△	
○	
○	
○	
○	

#### 2.2. Solvent degassing unit

2.2.1. Include a degassing device capable of eliminating gas generated during solvent blending for both isocratic and gradient elution.

2.2.2. The unit should include at least 2 solvent channels.

2.2.3. Each degassing channel shall be able to accommodate a flow rate from 0.001 mL/min to 5 mL/min or wider.

○	
○	
○	
○	

#### 2.3. Automated sample injection unit

2.3.1. The unit shall be compatible with the liquid chromatograph and system control unit for fully automatic operation.

2.3.2. The unit should include automated injection syringe with an appropriate needle and a sample loop. ( Support 2 ml, 105 and 96-well microtiter plate: 192)

2.3.3. Injection cycle time shall not be greater than 30 sec using default setting.

○	
○	
○	

2.3.4.	Injection range should be 0.1 to 50 $\mu$ l or wider. In 0.1 $\mu$ L increment without loop replacement or other hardware modification.	<input type="radio"/>	
2.3.5.	Injection precision should be <0.25% RSD or SD <10 nl.	<input type="radio"/>	
2.3.6.	Injection volume accuracy should be $\pm$ 1% (50 $\mu$ L, n = 10) max.	<input type="radio"/>	
2.3.7.	Carryover: <0.003% without washing, and <0.005% at 30 ppm or better.	<input type="radio"/>	
2.3.8.	Internal and external needle wash shall be available.	<input type="radio"/>	
2.3.9.	Injection port back-flush shall be available to reduce carry over.	<input type="checkbox"/>	
2.3.10.	The unit shall be able to operate from 4°C to 40°C or wider.	<input type="radio"/>	
2.3.11.	The temperature accuracy shall not be greater than $\pm$ 2°C.	<input type="radio"/>	
2.3.12.	The temperature stability shall not be greater than $\pm$ 1°C.	<input type="checkbox"/>	

#### 2.4. Thermostatted column compartment

2.4.1.	The temperature range shall be 5°C to 80°C, or wider. The compartment can be pre-heated up to 150°C.	<input type="radio"/>	
2.4.2.	Temperature stability: $\pm$ 0.1°C.	<input type="radio"/>	
2.4.3.	Temperature accuracy: $\pm$ 0.5°C.	<input type="radio"/>	
2.4.4.	The unit shall be able to house at least four columns with 300 mm in length with pre-column. Corresponding column switching valve and drive for four columns switching shall be provided.	<input type="radio"/>	
2.4.5.	The unit shall be able to house at least eight columns with 100 mm in length with pre-column.	<input type="checkbox"/>	
2.4.6.	The system shall have leak detection and leak output signal for shutdown of pumping system.	<input type="radio"/>	

### 3. High-Resolution Mass Spectrometry System

3.1.	High performance accurate mass Q-TOF for toxicologist analysis. It shall include a mainframe with ion optics and electronics, vacuum system and the mass analyzer. It shall be coupled to the HPLC system via an atmospheric pressure ionization interface for independent operation. The operation of the analyzer shall be controlled by the Workstation (section 4). The mass spectrometer shall be benchtop model or equivalent.	<input type="radio"/>	
3.2.		<input type="radio"/>	

#### 3.3. Ionization source

3.3.1.	All source parameters are under computer control except sprayer position.	<input type="radio"/>	
3.3.2.	The system should be equipped with separate ionization sources and capable to operate in electrospray ionization (ESI) source and an atmospheric pressure chemical ionization (APCI) source with both positive and negative ion mode.	<input type="radio"/>	
3.3.3.	Venting the instrument shall not be required when switching between ESI and APCI modes.	<input type="radio"/>	
3.3.4.	The ESI and the APCI source shall be interchangeable in the same housing.	<input type="checkbox"/>	
3.3.5.	The source should have orthogonal spraying for improved robustness. And should have self-cleaning ceramic heaters with embedded temperature sensor.	<input type="checkbox"/>	
3.3.6.	The source should have flow rate compatibility in ESI mode from 5 $\mu$ l/min to 3,000 $\mu$ l/min, and APCI mode from 50 $\mu$ l/min to 3,000 $\mu$ l/min without flow splitting.	<input type="radio"/>	
3.3.7.	The solvent compatibility is from 100 % aqueous to 100 % organic at flow rate up to 1 mL/min for ESI and APCI mode.	<input type="radio"/>	
3.3.8.	The desolvation temperature is user selectable from ambient to 750°C.	<input type="radio"/>	
3.3.9.	The nebulized spray of the ion source shall be positioned off-axis to reduce contamination of the mass spectrometer.	<input type="radio"/>	
3.3.10.	The ESI and APCI probes should be with two inlet channels for separate delivery of calibrant and sample solutions to the ionization chamber.	<input type="radio"/>	

#### 3.4. High-Resolution Mass Spectrometer (HRMS)

3.4.1.	The mass spectrometer shall include a quadrupole mass filter, a collision cell for MS/MS fragmentation and a second mass analyzer for accurate mass measurement.	<input type="radio"/>	
3.4.2.	The full scan mass range shall be 50-40,000 m/z or wider	<input type="checkbox"/>	

3.4.3.	The quadrupole mass range for precursor selection shall be 50-2200 m/z or wider.	<input type="radio"/>	
3.4.4.	The system shall include a calibrant delivery system (CDS) to deliver calibration standards via a separate ionization probe that functions independently from the main	<input type="radio"/>	
3.4.5.	The system shall be capable of optimizing quadrupole parameters automatically.	<input type="radio"/>	
3.4.6.	The system shall be capable of calibration mass accuracy and resolution automatically.	<input type="radio"/>	
3.4.7.	The mass spectrometer shall be able to, but not limited to, perform the functions as stipulated below.		
	(a) Full ion scan MS	<input type="radio"/>	
	(b) Q1 Scan	<input type="radio"/>	
	(c) MS/MS	<input type="radio"/>	
	(d) Product ion scan	<input type="checkbox"/>	
3.4.8.	Able to perform Data Dependent Acquisition (DDA), Information Dependent Acquisition (IDA) or equivalent.		
	(a) Able to perform neutral loss triggered DDA/IDA	<input type="checkbox"/>	
	(b) Able to perform DDA/IDA in precursor isolation window	<input type="radio"/>	
	(c) The collision energy applied to DDA/IDA scan shall be user controllable.	<input type="radio"/>	
	(d) Spectral library matching shall be feasible with the MS/MS spectrum acquired in DDA/IDA scan.	<input type="radio"/>	
3.4.9.	Able to perform Data Independent Acquisition (DIA), Information Independent Acquisition (IIA) or equivalent.		
	(a) The system shall acquire a high resolution, accurate mass MS/MS full spectrum in a specified Q1 window in DIA/IIA scan.	<input type="radio"/>	
	(b) Shall be able to acquire by fragmenting all ions that enter the mass spectrometer with high-resolution, accurate mass detection.	<input type="radio"/>	
	(c) Spectral library matching shall be feasible with the MS/MS spectrum acquired in DIA/IIA scan.	<input type="radio"/>	
	(d) The system shall acquired up to 200 Q1 windows per cycle during DIA/IIA scan.	<input type="checkbox"/>	
	(e) The collision energy applied to DDA/IDA and DIA/IIA scan shall be user controllable.	<input type="radio"/>	
	(f) The tandem mass spectra could be acquired by sequentially isolating and fragmenting ranges of m/z.	<input type="checkbox"/>	
3.4.10.	Able to quantitate using MS/MS data.	<input type="radio"/>	
3.4.11.	Sensitivity of exact mass measurement: signal to noise (S/N) for 1pg reserpine in ESI mode on column with positive ionization shall be 750:1 (RWS) or better.	<input type="checkbox"/>	
3.4.12.	MS/MS sensitivity: 1pg reserpine in ESI mode on column with positive ionization should produce S/N>1,500:1 RMS.	<input type="checkbox"/>	
3.4.13.	Scan speed:		
	(a) MS: at least 25Hz	<input type="radio"/>	
	(b) MS/MS: at least 100Hz	<input type="checkbox"/>	
3.4.14.	Mass resolution should be equal to or greater than 42,000 (FWHM).	<input type="radio"/>	
3.4.15.	Internal mass accuracy should be equal or less than 1 ppm.	<input type="radio"/>	
3.4.16.	Internal mass accuracy should be equal or less than 0.5 ppm.	<input type="checkbox"/>	
3.4.17.	Mass accuracy over time shall be less than 2ppm RMS over 12 hours or better.	<input type="checkbox"/>	
3.4.18.	Dynamic range should be equal to or greater than 4 orders of magnitude.	<input type="radio"/>	
3.4.19.	Computer controlled polarity switching shall be available.	<input type="radio"/>	
3.4.20.	The system shall include dynamic background subtraction technology to minimize collection of background ions during MS/MS analysis.	<input type="checkbox"/>	
3.5.	<b>Vacuum system</b>		
3.5.1.	The mass spectrometer shall include a differentially pumped vacuum system or equivalent.	<input type="radio"/>	
3.5.2.	The pumping vacuum system should preferably have fail-safe protection in case of	<input type="radio"/>	

- 3.5.3. Vacuum read-backs and system vent/pump cycles shall be digitally monitored to provide software control. 

<input type="radio"/>	
-----------------------	--
- 3.5.4. Oil and mist traps for the roughing pumps, if required, shall be provided. 

<input type="radio"/>	
-----------------------	--
- 3.5.5. The exhaust waste generated from the vacuum venting system shall be connected with the existing venting system at no extra 

<input type="radio"/>	
-----------------------	--

**4. System control, Data Management and Data Processing Unit**

<input type="radio"/>	
-----------------------	--

4.1. The workstation shall include, but not limited to the following units:

- (a) The system control unit (online workstation, section 4.2) for controlling the operation and handling of data. 

<input type="radio"/>	
-----------------------	--
- (b) A set of standalone offline data processing unit (offline workstation, section 4.3) 

<input type="radio"/>	
-----------------------	--

**4.2. System control and data handling unit (online workstation)**

4.2.1. All the interfaces and software necessary for the control and data processing of the HPLC-HRMS system shall be provided. 

<input type="radio"/>	
-----------------------	--

4.2.2. The online workstation shall be networked to the offline data processing unit. The installation of the local area network shall be provided without any additional cost. 

<input type="radio"/>	
-----------------------	--

4.2.3. The minimum hardware configuration of the online workstation is listed as follow:

- (a) Intel® core i5, quad core processor of speed 2.9 GHz 

<input type="radio"/>	
-----------------------	--
- (b) 32GB DDR3 RAM memory 

<input type="radio"/>	
-----------------------	--
- (c) Two sets of 2TB hard disk for storage of data 

<input type="radio"/>	
-----------------------	--
- (d) Three ethernet 10/100 Mbps LAN adapter cards with 3 ethernet ports 

<input type="radio"/>	
-----------------------	--
- (e) Four USB 3.0 ports 

<input type="radio"/>	
-----------------------	--
- (f) One sound card 

<input type="radio"/>	
-----------------------	--
- (g) Graphic card(s) with 64 MB memory or above, supporting dual monitor display 

<input type="radio"/>	
-----------------------	--
- (h) One 22-inch LED monitor or better 

<input type="radio"/>	
-----------------------	--
- (i) DVD-rewritable drive with 8X speed or above 

<input type="radio"/>	
-----------------------	--
- (j) One set of keyboard and mouse 

<input type="radio"/>	
-----------------------	--
- (k) One laser printer for automatic both sides printing with printing speed of up to 40 A4-pages per minute and up to 1200 x 1200 dpi resolution 

<input type="radio"/>	
-----------------------	--

4.2.4. The online workstation shall include:

- (a) A licensed system control and data handling software (section 4.2.5) 

<input type="radio"/>	
-----------------------	--
- (b) A licensed operating system such as Microsoft® Windows® 7 or its latest version, or equivalent. Vendor is responsible for any update required throughout the life of the analyzer. 

<input type="radio"/>	
-----------------------	--
- (c) Data processing software such as Microsoft® Office 2013 professional or higher version or equivalent with perpetual license for data processing and handling through database, spreadsheet, word processing and reporting. 

<input type="radio"/>	
-----------------------	--
- (d) Fully licensed anti-virus software such as Norton Anti-virus or equivalent software which is fully compatible with the workstation. 

<input type="radio"/>	
-----------------------	--

4.2.5. The system control and data handling software of the online workstation shall have the following features:

- (a) The software shall be able to operate and control the HPLC and HRMS system, enable instrument set-up, data acquisition, processing and reporting. 

<input type="radio"/>	
-----------------------	--
- (b) The software shall have true multi-tasking capacity to display, review and manipulate previously acquired data files while acquisition of new data is in progress. 

<input type="radio"/>	
-----------------------	--
- (c) The software shall be capable of displaying the complete timed event methods and real time display of instrument parameters and signals of the HPLC and HRMS systems. 

<input type="radio"/>	
-----------------------	--
- (d) Licensed software for generating system mirror image files for the workstation shall be provided for system recovery in case of computer break down. The system mirror image files shall contain the operating system, all licensed software, equipment settings and control parameters and acquired data. 

<input type="radio"/>	
-----------------------	--
- (e) All software shall come with official licenses. The licensed software shall be fully compatible with the operating system. 

<input type="radio"/>	
-----------------------	--

(f) The workstation software should be based on Microsoft® Windows® platform, compatible with Windows® 7 or higher version. Multitasking suite of analytical applications and instrument management software should be included for data processing and handling.	<input type="radio"/>	
(g) Complete data processing program including manual and automatic integration of chromatographic peaks/ noise threshold / baseline control parameters shall be provided.	<input type="radio"/>	
(h) The software shall be able to calculate the elemental composition. A goodness of fit from actual to theoretical isotopes or similar indicators shall be included. The software shall be able to filter out incorrect elemental composition calculations through the use of intelligent algorithms or other equivalent methodologies.	<input type="radio"/>	
(i) The software shall enable calibration using MS/MS data with linear or quadratic fits. Calibration information for each compound shall be accessible on screen, and shall include target and qualifier ions, retention time windows and regression outputs.	<input type="radio"/>	
(j) The software shall be capable of providing database searching function for matching of HR-MS and HR-MS/MS spectra.	<input type="radio"/>	
(k) Data analysis software for drug screening and quantitation shall be included.	<input type="radio"/>	
(l) The software shall be able to generate qualitative as well as quantitative reports. The report format shall be customized according to user's requirement at no extra cost.	<input type="radio"/>	
(m) Comprehensive self-diagnostic and extensive on-line help programs for fast troubleshooting of the whole system shall be provided.	<input type="radio"/>	
(n) Free upgrade of system software within 2 years after installation shall be provided.	<input type="radio"/>	

**4.3. A set of standalone offline data processing unit**

4.3.1. All the interfaces and software necessary for the control and data processing of the HPLC-HRMS system (section 2 to 3) shall be provided.	<input type="radio"/>	
4.3.2. The data processing unit shall be networked to the online workstation. The installation of the local area network shall be provided without any additional cost.	<input type="radio"/>	
4.3.3. The minimum hardware configuration of the online workstation is listed as follow:	<input type="radio"/>	
(a) Intel® Xeon Processor (20MB Cache, 3.2 GHz) or above	<input type="radio"/>	
(b) 32GB DDR3 RAM memory	<input type="radio"/>	
(c) Two sets of 2TB hard disk for storage of data	<input type="radio"/>	
(d) Three ethernet 10/100 Mbps LAN adapter cards with 3 ethernet ports	<input type="radio"/>	
(e) Four USB 3.0 ports	<input type="radio"/>	
(f) One sound card	<input type="radio"/>	
(g) Graphic card(s) with 64 MB memory or above, supporting dual monitor display	<input type="radio"/>	
(h) One 23-inch LED monitor or better	<input type="radio"/>	
(i) DVD-rewritable drive with 8X speed or above	<input type="radio"/>	
(j) One set of keyboard and mouse	<input type="radio"/>	
(k) One laser printer for automatic both sides printing with printing speed of up to 40 A4-pages per minute and up to 1200 x 1200 dpi resolution	<input type="radio"/>	
4.3.4. The offline data processing unit shall include:	<input type="radio"/>	
(a) A licensed system control and data handling software (section 4.2.5)	<input type="radio"/>	
(b) A licensed operating system such as Microsoft® Windows® 7 or its latest version, or equivalent. Vendor is responsible for any update required throughout the life of the	<input type="radio"/>	
(c) Data processing software such as Microsoft® Office 2013 professional or higher version or equivalent with perpetual license for data processing and handling through database, spreadsheet, word processing and reporting.	<input type="radio"/>	
(d) Fully licensed anti-virus software such as Norton Anti-virus or equivalent software which is fully compatible with the workstation.	<input type="radio"/>	

**5. Method package for clinica/forensic toxicology screening**

5.1. A validated data acquisition method of the high resolution mass spectrometer (HRMS) for forensic/clinical toxicology screening of urine and blood matrix shall be provided.	<input type="radio"/>	
--	-----------------------	--

- |   |  |
|---|--|
| ○ |  |
|---|--|
- 5.2. The method shall be compatible with the hardware and software described in section 2, 3 and
- 5.3. The details of mass spectrometric detection method shall include but not be limited to the
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 5.3.1. Details of chromatographic separation which include the composition and gradient of the mobile phase solutions, flow rate, injection volume, total run time, column type
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 5.3.2. Details of mass spectrometric parameter settings.
- 5.3.3. A forensic/clinical drug and metabolite database that contains high resolution MS/MS spectrum of least 600 drug and metabolites in positive and negative ionization mode. The information provided in the library shall include but not be limited to the retention time (in connection with the proposed chromatographic method), the accurate mass of molecular ions and fragmented ions. Fragment mass spectra of parent drugs shall be generated experimentally using authentic drug standards. Fragment mass spectra of metabolites are generated using authentic standards or biological samples with known clinical history in the presence of corresponding parent drugs.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 5.3.4. A designated result interpretation method and corresponding customized report format.
- 5.3.5. Additional database/library shall be provided (if available)
- |   |  |
|---|--|
| △ |  |
|---|--|
- (a) High resolution MS/MS library of pesticides
- |   |  |
|---|--|
| △ |  |
|---|--|
- (b) High resolution MS/MS library of natural products
- |   |  |
|---|--|
| △ |  |
|---|--|
- (c) High resolution MS/MS library of mycotoxins
- |   |  |
|---|--|
| △ |  |
|---|--|
- (d) High resolution MS/MS library of antibiotics
- |   |  |
|---|--|
| △ |  |
|---|--|

## 6. Nitrogen generation system

- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.1. An integrated nitrogen generation system, or a standalone nitrogen generator equipped with an air compressor, or other equivalent gas generator shall be provided to supply nitrogen gas for the operation of the HPLC-MS system.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.2. The nitrogen gas generators shall be fully compatible with the system. All accessories for installation such as tee(s), Teflon<sup>®</sup> connection tubing(s) required shall be provided.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.3. Nitrogen flow and purity shall be compatible with the MS operation.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.4. Noise level generated shall be lower than 60 dB within 10 meters during operation.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.5. The nitrogen gas generators shall be connected with the existing nitrogen gas back-up system at no extra cost.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.6. All necessary regulators and gas transfer lines for the connection between the instruments and gas supplies shall be provided.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 6.7. Any license required for the operation nitrogen generator or related shall be provided.
- |   |  |
|---|--|
| ○ |  |
|---|--|

## 7. Uninterrupted power supply (UPS)

- |   |  |
|---|--|
| ○ |  |
|---|--|
- 7.1. A constant and uninterrupted power supply (UPS) shall be assured to the LC-MS/MS (including HPLC, HRMS and workstations) for a minimum of 8 minutes of electricity backup time.
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 7.2. The UPS shall comply with IEC 62040, or equivalent standards, for electromagnetic compatibility requirements.

## 8. 配件、安裝及測試

- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.1. 必需為HPLC (section 2)、HRMS (section 3)及workstation system (section 4) 提供一個定制的工作枱，以便準確安裝儀器；
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.2. 必需包括一個以非易燃材料製成的通風隔音罩，以容納真空泵；
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.3. 工作台必需包括一減振設備，以避免primary vacuum pump對MS所造成的振動；
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.4. 應提供儀器的所有必要的校準和performance check kits；
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.5. 應提供用於儀器的日常及定期維護的專用工具。
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.6. 投標人應當確定電力需求，其環境溫度濕度，以優化儀器安裝的運作程序及其他特殊要求。
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.7. 儀器設備的電源操作應220V, 50Hz；以及適合用於IEC 60309 (30A) 之插頭。
- |   |  |
|---|--|
| ○ |  |
|---|--|
- 8.8. 在測試期間遇到問題時，應提供所有必要的協助，諮詢，時間和資源，以確保整個系統在整個系統的整個使用壽命期間完成安裝和平穩運行。
- |   |  |
|---|--|
| ○ |  |
|---|--|

**9. 培訓教學**

- 9.1. 儀器安裝完成後，需為實驗室人員提供兩次內容相同，且每次不少於14天的現場培訓。培訓應包括教授理論、軟件操作以及設備的使用和維護實踐課程。並應頒發培訓證書給已完成課程的學員。兩次的培訓時相距不少於6個月。
- 9.2. 應為實驗室之人員提供一額外的進階培訓 (on-site)。內容應包括動手系統操作、維護、實驗方法開發、故障排解以及軟件的應用程序培訓。
- 9.3. 第9.1及9.2節中所描述的培訓課程應由應用化學家或同等的專業人員以中文授課，並可以英文輔助進行，且無需額外付費。

○	
○	
○	

**10. 保養服務及技術支援**

- 10.1. 保用期內之保養服務
  - 10.1.1. 自設備驗收之日起計算(即在系統上完成用戶功能測試後)，投標人應對所提供的設備及所有物品提供至少一年的保養。在保養期內所有保養服務均應為免費，包括由合格的維修保養人員更換故障部件、維修計劃和故障維修服務；
  - 10.1.2. 在澳門特別行政區公共部門之辦公時間內，維修保用之電話接聽服務的回覆時間不得超過一小時；而維修人員應在收到電話後24小時內到達現場解決問題；
  - 10.1.3. 確保在保用期內（包括辦公日、非辦公日、假期等任何時間），儀器設備的停機時間不得超過連續三個工作日。獲判給人應存備有足夠數量的儀器設備基本配件及要件部件，並存放於其辦事處內。不接受任何以運輸配件的理由以作為延長儀器停機時間的原因；
- 10.2. 保用期後之保養服務
  - 10.2.1. 必需為設備提供在保用期後的維修保養服務，該維修保養服務時間為五年，並會以年度支付的方式作支付；
  - 10.2.2. 第10.2.1所指之維修保養服務內容需包括每年兩次的預防性維護及維修保養的工作；預防性維護應包括所有工作（當中必需包括檢查和調整），並能使整個設備（包括附件）達到有效性能規範，且承諾應備有足夠數量的備用零件部件，不能因所更換部件由於運輸原因而使儀器處於長期停止運作的狀態；
  - 10.2.3. 在收到故障通知後，在任何時間內，必須提供緊急服務（爭取24小時內，但不超過收到故障通知後連續三個工作日），價格應包括人工、交通費用及為設備進行預防性及故障性維修所需的全部零配件及物料。
  - 10.2.4. 應註明正常辦公時間所指之時段(以澳門特別行政區公共部門之辦公時間為標準)；及註明在非正常辦公時間內提供服務需收取的費用及其計算方式。
- 10.3. 應保證在儀器安裝後至少10年內，為儀器提供所需的零(配)件及要件以及支援服務。
- 10.4. 必須提供儀器設備所必要的備用零件、消耗品和附件的詳細價格及建議數量之清單，並說明其交貨時間。
- 10.5. 應在儀器設備的整個使用生命週期內，以具有經驗豐富及專業知識的化學家為儀器設備提供方法的開發和研究等專業建議。

○	
○	
○	
○	
○	
○	
○	
○	
○	
○	
○	
○	

投標人

(簽署及蓋章)  
二零一九年 月 日